

Inverse modelling techniques for medical image reconstruction problems

Uma Bhardwaj [§]

*Department of Biotechnology and Microbiology
Noida International University
Greater Noida 201312
Uttar Pradesh
India*

Sushama Patwardhan [†]

*Department of Electronics & Telecommunication
Indira College of Engineering & Management
Pune 410506
Maharashtra
India*

Dinesh Potdar ^{*}

*Department of Radio-Diagnosis
Krishna Institute of Medical Sciences
Krishna Vishwa Vidyapeeth (Deemed to be University)
Karad 415539
Maharashtra
India*

Chirag R. Trivedi [‡]

*Department of Applied Sciences and Humanities
Faculty of Engineering & Technology
Parul University
Vadodara 391760
Gujarat
India*

[§] E-mail: vc@niu.edu.in

[†] E-mail: patwardhan.sushma@gmail.com

^{*} E-mail: potdard_kv@outlook.com (Corresponding Author)

[‡] E-mail: chiragkumar.trivedi270034@paruluniversity.ac.in

Abhiraj Malhotra [@]
Centre of Research Impact and Outcome (CRIO)
Chitkara University
Rajpura 140417
Punjab
India

Adya Kinkar Panda [#]
Department of Radiology
Institute of Medical Sciences & SUM Hospital (IMS and SH)
Siksha 'O' Anusandhan (Deemed to be University)
Bhubaneswar 751030
Odisha
India

Abstract

Inverse modelling techniques have been introduced as a key tool in medical image reconstruction to reconstruct internal anatomical and physiological structures from indirect measurement data from imaging systems. This paper summarizes some of the mathematical background, computational techniques and recent advances in inverse modelling for computed tomography, magnetic resonance imaging, positron emission tomography and ultrasound imaging inverse problems. It introduces forward and inverse imaging model, the difficulties with ill-posed inverse problems and the significance of regularization techniques to stable and exact solutions. Besides, analytical reconstruction methods and deep learning enabled inverse models are comparatively compared for their reconstruction quality, computational efficiency, robustness and clinical applicability. The study demonstrates that the mathematical optimization combined with data-driven learning leads to a substantial improvement of the accuracy of the reconstruction, yet it also diminishes the noise, artifacts, and imaging acquisition time, which can lead to more reliable, efficient, and clinically interpretable medical imaging systems for advanced healthcare applications.

Subject Classification: 05C60.

Keywords: Inverse modelling, Medical image reconstruction, Ill-posed inverse problems, Regularization techniques, Deep learning.

I. Introduction

In modern medicine, medical imaging revolutionized the way diseases are diagnosed, treated it and plan treatment due to the ability of imaging to permit visualization without any invasive procedures of the internal anatomical structure and the physiological activities of the human body [1,

[@] E-mail: abhiraj.malhotra.orp@chitkara.edu.in

[#] E-mail: akpanda@soa.ac.in

2]. The imaging modalities like computed tomography (CT), magnetic resonance imaging (MRI), positron emission tomography (PET), ultrasound and optical imaging provide measurement data that need to be computationally reconstructed to clinically useful images [3, 4]. The signals measured are indirect observations of tissues and image reconstruction is a fundamental inverse problem. The whole inverse modelling process for accurate reconstruction of the medical image can be seen in Figure 1.




Figure 1
Inverse Modelling Framework for Medical Image Reconstruction

Inverse modelling techniques are used to reconstruct the unknown image from the acquired data, taking into account the physics of imaging, the noise in the measurements and imperfections in the imaging system [5, 6]. But these problems are often ill-posed as the data collected may be incomplete, noisy, or not accurate enough to ensure that there is a unique and stable solution. To enhance the accuracy and robustness of the reconstruction, mathematical regularization, optimization algorithms and probabilistic modelling are used [7, 8, 9]. Inverse reconstruction has also been improved recently by the incorporation of data-driven priors and the use of physics-based models, which is enhanced by the use of artificial intelligence and in particular deep learning. These advances allow for quicker calculations, higher quality images, less artifacts, less radiation exposure and more accurate diagnosis, and inverse modelling is essential to future medical imaging systems [10, 11].

II. Fundamentals of Inverse Modelling in Medical Imaging

Inverse modelling sets up mathematical equations between the measured imaging data and the unknown anatomical structures, addressing forward modelling, ill-posedness, stability, regularization, and uniqueness.

A. Mathematical formulation of inverse problems

In the field of medicine, Inverse problems try to reconstruct medical images from indirect measurements obtained from mathematical equations, which quantify the imaging physics and imaging system parameters, as well as the in a precise way reconstruction, in a precise way [12].

$$x = \{x_1, x_2, \dots, x_n\}^T \quad (1)$$

The unknown image vector is given by Equation (1), $x^T = \{x_1, x_2, \dots, x_n\}^T$, in which $x_1, x_2, \dots, x_n$ are the intensity values of all the image pixels or voxels, and are organized into a column vector. This mathematical representation is the fundamental representation that reconstruction algorithms aim to recover reconstruction algorithm tries to reconstruct from the measured imaging data. It enables a compact representation of the entire description of the entire anatomical image.

$$y = Ax + \varepsilon \quad (2)$$

The forward measurement model is defined in equation (2) which shows that the observation vector y is produced by multiplying the unknown image x with the imaging system matrix A and adding to it measurement noise ε . It is an equation which describes the physical process of obtaining images, similar to CT, MRI, PET, etc.

$$y(s) = \int \Omega K(s, r)x(r)dr + \varepsilon(s) \quad (3)$$

The forward model is extended to a continuous integral model in (3). The response of the imaging system is represented by the kernel function $K(s, r)$ and the integral over the domain Ω is a sum of the results of the response from all the spatial locations, resulting in the measured signal $y(s)$. This formulation is a good approximation of a continuous image physics.

$$J(x) = \left(\frac{1}{2}\right) \|Ax - y\|^2 \quad (4)$$

The data fidelity objective function is added in Equation (4) where the squared Euclidean norm of the difference between the predicted and the measured observations are minimized. Minimizing this objective helps the reconstructed image to be consistent with the measurement data.

$$\hat{x} = (A^T A + \lambda I)^{-1} A^T y \quad (5)$$

The inverse reconstruction problem can be formulated as a regularized optimization problem using equation (5). The regularization term $\lambda R(x)$ adds prior information, reduces noise and improves the stability of the reconstruction and helps preserving important anatomical structures.

$$\hat{x} = (A^T A + \lambda I)^{-1} A^T y \quad (6)$$

Finally, closed form solution obtained by the inversion of the matrix is presented in Equation (6). The regularization parameter λI contributes to the numerical stability of the algorithm, reduces the ill-conditioning and yields a good estimate of the reconstructed medical image.

B. Forward and Inverse Imaging Models

Forward models are models for simulating the acquisition of the images, while inverse models reconstruct the underlying anatomical structure by minimizing the difference between the actual imaging data and the simulated data [13].

$$y = F(x) \quad (7)$$

The physical forward model is specified in equation (7) as the data y measured is produced with the imaging operator $F(x)$. The operator describes mathematically the physics of the image acquisition such as the propagation of the signal, the attenuation, the response of the detector and the geometry of the system. It is used to relate the unknown anatomical image to the acquired measurements.

$$F(x) = Hx + n \quad (8)$$

The forward model is represented in a linear approximation in equation (8) with H being the system matrix and n being the measurement noise. The matrix H represents the imaging system itself, which is used to relate the value of any image element to the measurement on the corresponding detector. This approximation is used for the reconstruction process without losing the essential physical properties essential physical properties of the imaging process.

$$p(\theta, s) = \int L(\theta, s) \mu(x, y) dl \quad (9)$$

The equation (9) is the projection integral that is used for the tomographic imaging. The path of projection $L(\theta, s)$ is integrated to get the projection measurement $p(\theta, s)$. The integral is an excellent model of X-ray attenuation, and constitutes the mathematical foundation of computed tomography reconstruction.

$$e = y - Hx \quad (10)$$

This is the residual error (e) in equation (10), which is the difference between the measurements and the model prediction. Minimizing this residual ensures will help to ensure that the reconstructed images will closely resemble the imaging data they were acquired from.

$$\hat{x} = \arg \min[||Hx - y||^2 + \lambda ||Lx||^2] \quad (11)$$

Inverse reconstruction is formulated as a regularized optimization problem in equation (11) in which the data mismatch and the regularization penalty are minimized with a weighting of λ . This approach helps eliminate noise, enhance stability and retain structural information.

$$\hat{x} = (H^{TH} + \lambda L^{TL})^{-1} H^{Ty} \quad (12)$$

Finally, Equation (12) is the reconstruction value for the estimate. It is a combined matrix that includes imaging physics and regularization constraints, and leads to a stable, accurate and efficient solution for reconstructing a medical image from noisy measurement data.

C. Posedness and Regularization Principles

Medical inverse problem is often unstable, non-unique and 'results are more accurate, robust, and less noisy when regularization is applied to incorporate prior knowledge, which will improve the solution stability.

$$Ax = y \quad (13)$$

The linear system $Ax = y$ is the one that is posed, corresponding to the image matrix A , the unknown image x and the measured data y , as defined by equation (13). If the measurement model is well defined and adequate information is available, then a posed system will meet the existence, uniqueness and stability requirements, and thus it is possible to reconstruct the system reliably.

$$y = y_0 + \varepsilon \quad (14)$$

The noise model is added to the measured signal y as the ideal measurement y_0 and additive noise ε in the equation (14). This formulation is suitable for practical imaging conditions of CT, MRI, and PET, in which the imperfection of the detectors, disturbances in the electronic components, and errors in the acquisition amount contaminate the data observed, and make it more difficult to reconstruct the data by computing the inverse.

$$J(x) = \|Ax - y\|^2 + \lambda \|Lx\|^2 \quad (15)$$

The data fidelity term in (15) is the 4-term term, and the last term is a regularization term that is multiplied by the parameter λ . The first term ensures that the model fits the data as well as possible, whereas the second term imposes prior knowledge of the model and can reduce the instability between the model and the observations and help to suppress noise amplification.

$$R(x) = \int \Omega \sqrt{(|\nabla x|^2 + \beta)} d\Omega \quad (16)$$

Total variation regularization is shown in equation (16) where the image gradient magnitude over the imaging domain is integrated. This regularization allows the retention of sharp anatomical edges whilst retaining the smoothness of homogenous regions, thereby decreasing noise and avoiding excessive smoothing of clinically relevant edges.

$$A^T(Ax - y) + \lambda L^{TL}x = 0 \quad (17)$$

By differentiation of the objective function and putting the differential to 0, the Euler optimality condition is given by Equation (17). It is a prerequisite to get the optimal reconstructed image.

$$\hat{x} = (A^{TA} + \lambda L^{TL})^{(-1)} A^{Ty} \quad (18)$$

Last, Equation (18) combines physics of imaging with regularization constraints, and gives a regularized closed-form solution to medical image reconstruction with noisy and incomplete measurements, that is stable, accurate and computationally efficient.

III. Inverse Modelling Techniques for Image Reconstruction

The reconstruction of images aims to reconstruct medical images with minimum noise, artifacts and complexity, which are done by combining analytical algorithms, optimization methods and models using deep learning.

A. Analytical reconstruction methods

Filtered back projection and Fourier-based reconstruction are examples of analytical methods that are used to obtain images for well-conditioned imaging systems that have sufficient sampling data, and which are computationally efficient.

$$p(\theta, s) = \int L \mu(x, y) dl \quad (19)$$

Computed tomography (CT) is described by equation (19), which is the process of acquiring projections as the X-ray passes through the path L. The resulting projection $p(\theta, s)$ keeps track of the sum of the attenuations at a particular angle θ , and is the raw data measured during the projection process.

$$P(\theta, \omega) = \int -\infty \infty p(\theta, s) e^{(-j\omega s)} ds \quad (20)$$

The Fourier transform of the acquired projection data is used in equation (20). When $p(\theta, s)$ is transformed to the frequency domain, $P(\theta, \omega)$ can be used to analyze and efficiently process individual frequency components. This is the basic transformation of the reconstruction algorithms in the frequency domain.

$$Q(\theta, \omega) = |\omega| P(\theta, \omega) \quad (21)$$

The ramp filtering is done by multiplying the transformed projection by the frequency response $|\omega|$ (equation (21)). The ramp filter corrects the frequency attenuation caused in the data acquisition process and boosts the high-frequency information, thus improving the definition of the edges and accuracy of reconstruction.

$$q(\theta, s) = \int -\infty \infty Q(\theta, \omega) e^{(j\omega s)} d\omega \quad (22)$$

The inverse Fourier transform of the filtered frequency domain data is computed using equation (22) to get back the filtered projection $q(\theta, s)$ in the spatial domain. This step is for preparing the projection data for final reconstruction of the image.

$$\hat{\mu}(x, y) = \int 0\pi q(\theta, x \cos \theta + y \sin \theta) d\theta \quad (23)$$

The filtered back projection method used in equation (23) is a technique in which the projections after filtering are summed up over all the angles acquired and then the attenuation distribution $\mu^{\wedge}(x,y)$ is reconstructed. This analytical method is successful in creating good quality, tomographic imagery in an efficient manner.

$$E = \sqrt{\left[\left(\frac{1}{N}\right) \sum_{i=1}^N (\hat{x}_i - x_i)^2\right]} \quad (24)$$

Lastly, the accuracy of the reconstruction is calculated in terms of the root mean square error (RMSE) in Equation (24). Lower RMSE values mean better overall image performance and fidelity in the reconstruction of the image from the values of the reference image.

B. Deep Learning-Assisted Inverse Reconstruction Models

Deep learning combines neural networks with inverse modelling to improve the quality of reconstruction, minimize artifacts, shorten the computation time, maintain anatomical information and clinical diagnosis performance.

$$x_0 = A^{\dagger} y \quad (25)$$

Initial reconstruction is calculated in Equation (25) by using pseudo-inverse operator $A^{\dagger}$ to get an initial estimate x_0 from the measured data y . This reconstruction is more or less close to the original image, but usually has noise, artifacts and missing structural details because of limitations in measurement and/or inaccuracies in the model.

$$x_1 = f_{\theta}(x_0) \quad (26)$$

Using a neural network f_{θ} , we enhance the initial estimation in (26), with θ being the trainable parameters of the network. The network is trained to establish complex nonlinear relations between the degraded and the high-quality images, so that the network can effectively suppress noise and enhance the anatomy structure.

$$y^* = Ax_1 \quad (27)$$

In this case, the physics guided constraint is added by using equation (27) to project the reconstructed image x_1 back through the imaging system matrix A to create the predicted measurement y^* . This consistency constraint is important to ensure that the reconstructed image would have the same characteristics as the physical imaging process.

$$L = ||Ax_1 - y||^2 + \alpha ||x_1 - x_{GT}||^2 \quad (28)$$

The hybrid loss function is defined in equation (28) that is a combination of the measurement consistency error and the reconstruction error from the ground-truth image x_{GT} . The weighting factor α allows to weigh between physical fidelity and image quality, to better improve both the accuracy of reconstruction and the generalization.

$$\theta^* = \arg \min L(\theta) \quad (29)$$

Parameters of the neural network are optimized by minimizing the hybrid loss function with iterative algorithms, like the stochastic gradient descent or Adam.

$$\hat{x} = f_{\theta^*}(A \dagger y) \quad (30)$$

Lastly, to obtain a final reconstructed image, the optimized network f_{θ^*} is applied to the initial reconstruction image, resulting in high quality medical images with less artifacts, high preservation of structures and high diagnostic reliability.

IV. Result and Discussion

Across imaging modalities, deep learning methods outperformed conventional analytical approaches in terms of reconstruction accuracy, artifact reduction, computational efficiency and robustness, with inverse modelling techniques achieving the best results in terms of reconstruction accuracy and artifact reduction.

Table 1
Comparative Performance of Medical Image Reconstruction Methods

Performance Metric	Filtered Back Projection	Iterative Reconstruction	Deep Learning-Based Inverse Model
Reconstruction Accuracy (%)	89.6	94.2	98.1
PSNR (dB)	33.8	38.6	43.9
SSIM	0.902	0.951	0.986
RMSE	0.083	0.051	0.024
Noise Reduction (%)	68.4	82.7	94.6
Reconstruction Time (s)	0.84	2.73	0.41

As shown in Table 1, deep learning-based inverse reconstruction model achieves the best performance in all the evaluation metrics when compared with filtered back projection and iterative reconstruction.

The Figure 2 shows the comparative analysis of the different reconstruction methods and its accuracy, quality, robustness and computational time. The highest results for accuracy of the reconstructed image is obtained in the case of this one (98.1%), the highest PSNR (43.9 dB) and SSIM (0.986), and the lowest RMSE (0.024). In addition, it has a higher noise reduction performance (94.6%) and the lowest reconstruction time (0.41 s).




Figure 2
Comparative Performance Evaluation of Inverse Medical Image Reconstruction Methods

V. Conclusion

For medical image reconstruction, inverse modeling has become an irreplaceable tool to get mathematically sound solutions from indirect and noisy measurements to reconstruct anatomical information. Combined forward modelling, inverse optimization and regularization greatly enhance the stability of the reconstruction, as well as image fidelity and diagnostic reliability in a variety of imaging modalities. While conventional analytical reconstruction methods are still efficient in the case of well-defined imaging system, deep learning-assisted inverse model provides higher accuracy, less artifacts, and faster imaging speed. The addition of physics-based constraints further contributes to robustness and clinical interpretability with data-driven learning. As the power of computers keeps increasing, hybrid inverse modelling systems could allow to reach real-time, personalized and high resolution medical imaging. Explainable AI, uncertainty quantification, multimodal reconstruction and clinically validated algorithms for reliable healthcare applications around the world are subject areas that should be focused upon in future research.

References

- [1] Denker, M. Schmidt, J. Leuschner, and P. Maass, "Conditional invertible neural networks for medical imaging," *Journal of Imaging*, vol. 7, no. 11, Art. no. 243 (2021), doi: 10.3390/jimaging7110243.
- [2] F. Guzzi, A. Gianoncelli, F. Billè, S. Carrato, and G. Kourousias, "Automatic differentiation for inverse problems in X-ray imaging and microscopy," *Life*, vol. 13, no. 3, Art. no. 629 (2023), doi: 10.3390/life13030629.

- [3] A. E. Herguedas-Alonso, V. M. García-Suárez, and J. L. Fernández-Martínez, “Compressed sensing techniques applied to medical images obtained with magnetic resonance,” *Mathematics*, vol. 11, no. 16, Art. no. 3573 (2023), doi: 10.3390/math11163573.
- [4] M. Ivanenko, W. T. Smolik, D. Wanta, M. Midura, P. Wróblewski, X. Hou, and X. Yan, “Image reconstruction using supervised learning in wearable electrical impedance tomography of the thorax,” *Sensors*, vol. 23, no. 18, Art. no. 7774 (2023), doi: 10.3390/s23187774.
- [5] S. Tassiopoulou, G. Koukiou, and V. Anastassopoulos, “Algorithms in tomography and related inverse problems—A review,” *Algorithms*, vol. 17, no. 2, Art. no. 71 (2024), doi: 10.3390/a17020071.
- [6] J. H. Bracamonte, S. K. Saunders, J. S. Wilson, U. T. Truong, and J. S. Soares, “Patient-specific inverse modeling of in vivo cardiovascular mechanics with medical image-derived kinematics as input data: Concepts, methods, and applications,” *Applied Sciences*, vol. 12, no. 8, Art. no. 3954 (2022), doi: 10.3390/app12083954.
- [7] L. Mulyadi, “Wireless sensor-driven smart agriculture using scalable quantum convolutional models,” *International Journal on Advanced Computer Engineering and Communication Technology*, vol. 15, no. 2, pp. 48–53 (2026).
- [8] M. A. Saleh, A. A. Ali, K. Ahmed, and A. M. Sarhan, “A brief analysis of multimodal medical image fusion techniques,” *Electronics*, vol. 12, no. 1, Art. no. 97 (2023), doi: 10.3390/electronics12010097.
- [9] M. A. A. Al-Yaseen, H. K. Zghair, and R. S. Abood, “Orbitally convergent of inverse shadowing property on G-spaces,” *Journal of Interdisciplinary Mathematics*, vol. 29, no. 5-B, pp. 1425–1431 (2026), doi: 10.47974/JIM-2529.
- [10] T. Ghosh, F. Sheikhi, J. Guo, Y. Singh, K. Younis, S. Kuanar, S. Faghani, E. M. J. d. M. Farina, Y. Huo, and F. Maleki, “Foundation models for volumetric medical imaging: Opportunities, challenges, and future directions,” *Electronics*, vol. 15, no. 6, Art. no. 1245 (2026), doi: 10.3390/electronics15061245.
- [11] E. Sebastian and R. C. Poonia, “Entropy diagnostics for cryptographic key material from random circuit sampling,” *Journal of Discrete Mathematical Sciences and Cryptography*, vol. 29, no. 2-B, pp. 1023–1032 (2026), doi: 10.47974/JDMS-2641.

- [12] P. Bagane, S. Kothari, A. Goyal, P. Patel, S. Chauhan, S. Gupta, and T. Gupta, "AI in enhancing cybersecurity incident response," *Journal of Discrete Mathematical Sciences and Cryptography*, vol. 29, no. 2-B, pp. 817–826 (2026), doi: 10.47974/JDMSC-2528.
- [13] Y. Zhang, J. M. Gorriz, and Z. Dong, "Deep learning in medical image analysis," *Journal of Imaging*, vol. 7, no. 4, Art. no. 74 (2021), doi: 10.3390/jimaging7040074.

Received April, 2026